

Listing#: E115647  
Report #: 127567  
Original Certification Date: April 24, 2024



This Certification is issued to:  
Analytical Industries Inc.  
2855 Metropolitan Place  
Pomona, CA 91767 USA

Stating that the product(s):  
Oxygen Analyzer  
Models GPR-1000, GPR-1100, GPR-1200, GPR-1200 MS2, GPR-2000

Product Rating(s):  
The Electrical Rating for all the Models (GPR-1000, GPR-1100, GPR-1200, GPR-1200 MS2, GPR-2000) are 9Vdc, 2A 18W from an external DC source, and Internally Powered 6V.

Achieved Certification to the following standard(s):  
UL 61010-1, Third Edition: Safety of Electrical Equipment For Measurement, Control, and Laboratory Use, Rev. July 19, 2019.



Hon Keung Ip  
Eurofins Electrical and Electronic Testing North America, Inc.

*All changes proposed in the previously identified product that affects the above information must be submitted to Eurofins for evaluation prior to implementation to assure continued NRTL Certification status. The covered product(s) shall be subject to follow-up inspections to ensure that the Certified product(s) are identical to the product sample evaluated by Eurofins E&E NA and that all responsibilities are being fulfilled as specified in the Applicants' Responsibility section of the Certification Report. The Applicant named above has been authorized Eurofins E&E NA to represent the product(s) listed in this record as "MET Certified" and to mark this/these product(s) according to the terms and conditions of the Eurofins E&E NA Applicant Contract, Listing Reports, and the applicable agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in this Certification program. This certification has been granted under a System 3 program as defined in ISO/IEC 17067.*

*Eurofins E&E North America, Inc. is accredited by OSHA and the Standards Council of Canada.*

**N  
R  
T  
L**